



FM Approvals
1151 Boston-Providence Turnpike
P.O. Box 9102 Norwood, MA 02062 USA
T: 781 762 4300 F: 781 762 9375 www.fmglobal.com

CERTIFICATE OF COMPLIANCE

HAZARDOUS (CLASSIFIED) LOCATION ELECTRICAL EQUIPMENT

This certificate is issued for the following equipment:

3001. Insta Trans Oxygen Transmitter.

IS / I / 1 / ABCD / T4 Ta=50°C – D-73226; NI / I / 2 / ABCD / T4 Ta=50°C

Equipment Ratings:

Intrinsically Safe for use in Class I, Division 1, Groups A, B, C and D when installed in accordance with drawing number D-73226; Nonincendive for use in Class I, Division 2, Groups A, B, C and D Hazardous (Classified) Locations

Approved for:

Teledyne Analytical Instruments
16830 Chestnut St.
City of Industry, CA 91749



This certifies that the equipment described has been found to comply with the following FM Approval Standards and other documents:

Class 3600	1998
Class 3610	1988
Class 3611	1999
Class 3810	1989

Original Project ID: 3007225


FM Approval Granted: March 16, 2001

Subsequent Revision Reports / Date FM Approval Amended

3014720

August 9, 2002

FM Global Technologies LLC



David W. Styracula
Technical Team Manager

8/12/02
Date

