

DECLARATION OF CONFORMITY

APPLICATION OF COUNCIL : 89/336/EEC
ATEX DIRECTIVE : 94/9/EC

STANDARDS TO WHICH : EN 60079-0:2006
CONFORMITY IS DECLARED : EN 60079-11:2007

MANUFACTURER'S NAME : Teledyne Analytical Instruments

MANUFACTURER'S ADDRESS : 16830 Chestnut Street
City of Industry, CA 91748
U.S.A.

TYPE OF EQUIPMENT : Portable Oxygen Analyzer

EQUIPMENT CLASS : ISM Class A Group 1

MODEL NUMBER : 311PC, 311TC & 311TCXL

I, the undersigned, hereby declare that the equipment specified above conforms to the referenced standards (s) per 89/336/EEC and 94/9/EC and have been type-approved by Baseefa Limited, Staden Lane, Buxton, Derbyshire SK17 9RZ, UK. Baseefa notified body identification number: 1180. Mark: Ex II 2 G. Code: EEx ib IIC T3.

SIGNATURE: *Roger Starlin*

FULL NAME: Roger Starlin

POSITION: QA Manager

Date: 10-29-09

PLACE: City of Industry, California

SIGNATURE: *Stephen Broy*

FULL NAME: Stephen Broy

POSITION: Director of Engineering

Date: 10-29-09

PLACE: City of Industry, California